

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: South Yorkshire Police

Serial Number: LW06301

Certificate Number: 5909

Date of Calibration: 09 March 2022

Next Calibration: 08 March 2023

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999892 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.3 ft	284.3 ft	488.7 ft

Speed ± 1 MPH

Actual	56 MPH	73 MPH	-40 MPH	-66 MPH	27 MPH
Indicated	56 MPH	73 MPH	-40 MPH	-66 MPH	27 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD