

CERTIFICATE OF CONFORMITY

Date:	23/11/2020	Certificate Number:	20 – 1007
Customer:	SOUTH YORKSHIRE POLICE		
Serial Number:	TC003667 / 449490TC		
Instrument:	LTI 20:20 TRUCAM		

Visual Inspection:	Pass
Power On Self Test:	Pass
Display Integrity:	Pass
Firmware Version Check:	Pass
Distance Checks:	Pass
Pulse Repetition Check (Hz):	200.3799 Hz
Scope Alignment Check:	Pass

Distance Check	0 mph Reading	Speed Simulation	Measured Speed (mph)	Result +/- 1 mph
25m	Pass	20mph	19	Pass
75m	Pass	30mph	29	Pass
100m	Pass	70mph	70	Pass
125m	Pass	100mph	99	Pass
150m	Pass	150mph	149	Pass

This instrument has been verified to comply with the Manufacturer's Specification. The calibration/verification was controlled using documented procedures and test equipment that has been certified by UKAS approved test facilities.

We certify any parts or components that may have been used during these procedures are genuine parts as supplied by the manufacturer and identical to those used in the device as Type Approved.

For Laser Tech UK Ltd



Dominic Sadler

Next Calibration Due:	23/11/2021
-----------------------	------------