CALIBRATION CERTIFICATE

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:

South Yorkshire Police

Serial Number:

LW06302

Certificate Number:

6837

Date of Calibration:

23 February 2023

Next Calibration:

22 February 2024

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 19.999923 MHz

Self Test Sequence

- Pass

External Controls

- Pass

- Pass

Aiming

Horizontal - Pass

Vertical

Eye Safety

- Pass

Range ± 0.5 ft

Actual

156.3 ft

356.60 m

Indicated

156.3 ft

356.7 ft

488.4 ft 488.8 ft

Speed ± 1 MPH

Actual

27 MPH

-31 MPH

44 MPH

-57 MPH

72 MPH

Indicated

27 MPH

-31 MPH

44 MPH

-57 MPH

72 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

Technical Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400

email:info@truvelouk.com

